

Schedule

Supertron Sensing Pte Ltd
Blk 16 Kallang Place
#07-04
Singapore 339156

Certificate No. : LA-2011-0485-C
Issue No. : 14
Date : 5 May 2026
Expiry of Certificate : 10 May 2030
Page : 1 of 6

FIELD OF TESTING : Calibration and Measurement

MEASURED QUANTITIES/ INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD OF CALIBRATION	CALIBRATION AND MEASUREMENT CAPABILITY (CMC *)
<p>1. Pressure Measuring Devices</p> <ul style="list-style-type: none"> i. Calibrator ii. Transducer / Transmitter iii. Digital Indicator <p><u>Range of Measurement</u></p> <ul style="list-style-type: none"> a. (5 to 23) psi abs b. (23 to 50) psi abs c. (50 to 150) psi abs d. (150 to 300) psi abs 	<p>In-house Procedure CP-H, Rev 13</p>	<p>0.005 psi abs</p> <p>0.013 psi abs</p> <p>0.025 psi abs</p> <p>0.040 psi abs</p>
<p>2. On-Site Pressure Measurement</p> <ul style="list-style-type: none"> i. Calibrator ii. Transducer / Transmitter iii. Digital Indicator <p><u>Range of Measurement</u></p> <ul style="list-style-type: none"> a. (10 to 50) psi abs b. (50 to 150) psi abs 	<p>In-House / Site Procedure CP-H, Rev 13</p>	<p>0.03 psi abs</p> <p>0.10 psi abs</p>

Schedule



Certificate No. : LA-2011-0485-C

Issue No. : 14

Date : 5 May 2026

Page : 2 of 6

MEASURED QUANTITIES/ INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD OF CALIBRATION	CALIBRATION AND MEASUREMENT CAPABILITY (CMC *)
<p>3. Humidity Measurement</p> <p>i. Dew/ Frost Point</p> <p><u>Range of Measurement</u></p> <p>a. -95 °C to -90 °C Frost Point</p> <p>b. -90 °C to -85 °C Frost Point</p> <p>c. -85 °C to -75 °C Frost Point</p> <p>d. -75 °C to -60 °C Frost Point</p> <p>e. -60 °C to -30 °C Frost Point</p> <p>f. -30 °C to 0°C Frost/ Dew Point</p> <p>g. 0 °C to +70 °C Frost/ Dew Point</p> <p>h. +70 °C to +95 °C Frost/ Dew Point</p> <p>ii. Relative Humidity (Chilled Mirror Hygrometer with air temperature probe)</p> <p>iii. Relative Humidity Sensor/ Instrument</p> <p>iv. Thermo-hygrometer</p> <p><u>Range of Measurement</u></p> <p>a. At -10 °C to 0 °C (5 to 98) % relative humidity</p> <p>b. At 0 °C to 25 °C (2 to 98) % relative humidity</p> <p>c. At 25 °C to 50 °C (1 to 98) % relative humidity</p> <p>d. At 50 °C to 70 °C (1 to 98) % relative humidity</p> <p>e. At 70 °C to 95 °C (1 to 98) % relative humidity</p>	<p>In-house Procedure CP-N, Rev 6 CP-M, Rev 5</p> <p>In-house Procedure CP-N, Rev 6</p> <p>Comparison with 2- Pressure 2-Temperature Humidity generator rh = measured value</p>	<p>0.90 °C</p> <p>0.60 °C</p> <p>0.35 °C</p> <p>0.20 °C</p> <p>0.13 °C</p> <p>0.05 °C</p> <p>0.07 °C</p> <p>0.09 °C</p> <p>Corresponding to above dew-point and temperature uncertainties</p> <p>0.1% + 0.011 · rh[#]</p> <p>0.1% + 0.008 · rh</p> <p>0.1% + 0.004 · rh</p> <p>0.1% + 0.005 · rh</p> <p>0.1% + 0.007 · rh</p> <p>[#] ratio multiply by rh e.g. 0.011 x rh</p>

Schedule



Certificate No. : LA-2011-0485-C

Issue No. : 14

Date : 5 May 2026

Page : 3 of 6

MEASURED QUANTITIES/ INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD OF CALIBRATION	CALIBRATION AND MEASUREMENT CAPABILITY (CMC *)
<p><u>Range of Measurement</u></p> <p>a. At -10 °C to 0 °C (5 to 98) % relative humidity</p> <p>b. At 0 °C to 25 °C (2 to 98) % relative humidity</p> <p>c. At 25 °C to 70 °C (1 to 98) % relative humidity</p> <p>d. At 70 °C to 95 °C (1 to 98) % relative humidity</p>	<p>Comparison with a chilled mirror hygrometer with a temperature probe rh = measured value</p>	<p>0.16% + 0.008 · rh</p> <p>0.18% + 0.004 · rh</p> <p>0.21% + 0.003 · rh</p> <p>0.12% + 0.005 · rh</p>
<p>4. Lab/ On-Site Humidity Measurement</p> <p>i. Relative humidity generator (-10 to 70) °C</p> <p><u>Range of Measurement</u></p> <p>a. At -10 °C to 0 °C (10 to 95) % relative humidity</p> <p>b. At 0 °C to 23 °C (10 to 98) % relative humidity</p> <p>c. At 23 °C to 50 °C (10 to 98) % relative humidity</p> <p>d. At 50 °C to 70 °C (10 to 98) % relative humidity</p>	<p>In-house / Site Procedure CP-N, Rev 6</p> <p>Comparison with a chilled mirror hygrometer with a temperature probe rh = measured value</p>	<p>0.08 °C</p> <p>0.16% + 0.008 · rh</p> <p>0.18% + 0.004 · rh</p> <p>0.21% + 0.003 · rh</p> <p>0.15% + 0.003 · rh</p>
<p>ii. Relative Humidity Sensor / Instrument</p> <p>iii. Thermo-hygrometer</p> <p>iv. Hygrometers</p> <p>v. Temperature sensors incorporated in humidity instruments</p> <p>(0 to 25) °C</p> <p>(45 to 60) °C</p>	<p>In-house / Site Procedure CP-N, Rev 6</p>	<p>0.21 °C</p> <p>0.23 °C</p>

Schedule

Certificate No. : LA-2011-0485-C

Issue No. : 14

Date : 5 May 2026

Page : 4 of 6

MEASURED QUANTITIES/ INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD OF CALIBRATION	CALIBRATION AND MEASUREMENT CAPABILITY (CMC *)
<p><u>Range of Measurement</u></p> <p>a. At 0 °C to 25 °C (10 to 85) % relative humidity</p> <p>b. At 25 °C to 45 °C (10 to 95) % relative humidity</p> <p>c. At 45 °C to 60 °C (10 to 90) % relative humidity</p>	<p>Comparison with a chilled mirror hygrometer with air temperature probe rh = measured value</p>	<p>0.25% + 0.017 · rh</p> <p>0.20% + 0.013 · rh</p> <p>0.18 % + 0.012 · rh</p>
<p>5. Temperature</p> <p>A. Liquid bath method</p> <p>i. Temperature indicator and Recorders, with temperature sensor(s)</p> <p>ii. Industrial Platinum Resistance Thermometer</p> <p>iii. Thermistor</p> <p><u>Range of Measurement</u></p> <p>a. -196 °C</p> <p>b. (-100 to +5) °C</p> <p>c. (+5 to +250) °C</p> <p>d. Ice Point, 0 °C</p> <p>e. (0 to +30) °C</p> <p>f. Water Triple Point</p> <p>g. Gallium Fixed Point</p>	<p>In-house Procedure CP-C, Rev 4</p> <p>} Either by Fixed Point or } by comparison with } SPRT in liquid bath</p>	<p>20 mK</p> <p>20 mK</p> <p>20 mK</p> <p>10 mK</p> <p>10 mK</p> <p>10 mK</p> <p>10 mK</p>
<p>B. Air chamber method</p> <p>i. Temperature sensors incorporated in humidity instruments</p> <p>ii. Thermistor</p> <p>iii. Chilled Mirror Hygrometer with air temperature probe</p> <p>iv. Temperature indicator with temperature sensor</p>	<p>In-house Procedure CP-N, Rev 6</p>	

Schedule



Certificate No. : LA-2011-0485-C

Issue No. : 14

Date : 5 May 2026

Page : 6 of 6

MEASURED QUANTITIES/ INSTRUMENTS / RANGE TO BE CALIBRATED	METHOD OF CALIBRATION	CALIBRATION AND MEASUREMENT CAPABILITY (CMC *)
<p><u>Relative Humidity</u></p> <p>i. Climatic Chamber (Wet Mode)</p> <p>a. At 5 °C to 95 °C (1 to 98) % relative humidity</p> <p>ii. Climatic Chamber (Wet Mode) (Over reduced volume)</p> <p>a. At 5 °C to 95 °C (1 to 98) % relative humidity</p>	<p>Comparison with a chilled mirror hygrometer with air temperature probe rh = measured value</p>	<p>1.30 °C 0.23% + 0.027 · rh</p> <p>0.09 °C 0.17% + 0.013 · rh</p>

* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95%.

Approved signatories:

Mr Tee Yee Chee @Mr Zheng Yiqi - All items

Ms Eva Marie Barrera - All items

Note :

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid test results. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.